



## Manufacturer's Declaration to Regulation (EU) 2023/607

According to Regulation (EU)2017/745 (MDR) and regarding the transitional provisions Kuraray Noritake Dental Inc. (Miyoshi) declares the amendments to Article 120 of the MDR, as amended by Regulation (EU) 2023/607 applies to the following device(s):

Manufacturer name:	Kuraray Noritake Dental Inc.
Manufacturer address and contact details:	300 Higashiyama, Miyoshi-cho, Miyoshi, Aichi, 470-0293 Japan
Single Registration Number (SRN) (if available):	JP-MF-000017328

Authorised Representative name	Emergo Europe B.V.
Authorised Representative address and contact details	Westervoortsedijk 60, 6827 AT, Arnhem THE NETHERLANDS
Single Registration Number (SRN)	NL-AR-000000116

Notified body name (if applicable)	SGS Belgium NV <input checked="" type="checkbox"/> See attached schedule
Notified body number (if applicable)	1639 <input checked="" type="checkbox"/> See attached schedule
Directive Certificate number(s) to which this confirmation is made (if applicable)	JP19/040488 <input checked="" type="checkbox"/> See attached schedule
Original expiry date as indicated on the Directive Certificate prior to the extension of the validity (if applicable)	2023-10-24 <input checked="" type="checkbox"/> See attached schedule
End date of extended validity/transition period	2028-12-31 <input checked="" type="checkbox"/> See attached schedule

We, as the manufacturer declare under our sole responsibility:

- for the above listed **Directive Certificate** (or see attached schedule, if multiple certificates) the conditions for the legal extension of validity as required in Article 120.2 of the MDR are met *and/or*<sup>1</sup>
- the listed **device(s)** in the attached schedule and we as their manufacturer are in compliance with the conditions listed in Article 120.3c of the MDR for continued placing on the market and putting into service,

namely by fulfilling the following conditions:

➤ **Directive Certificate(s)** as listed above or in the attached schedule

- Directive Certificate(s) covering the listed device(s) was/were issued after 25 May 2017, was/were valid on 26 May 2021, was/were not withdrawn by 20 March 2023
- *Choose applicable statements:*
  - Expired *before* 20 March 2023:
    - Before the original date of expiry as indicated on the Directive Certificate, we and the notified body have signed a written agreement in accordance with Section 4.3, second subparagraph of Annex VII to this Regulation for the conformity assessment in respect of the device covered by the expired certificate or in respect of a device intended to substitute that device
    - A Competent Authority has granted a derogation from the applicable conformity assessment procedure in accordance with Article 59(1) MDR (may be provided upon request)
    - A Competent Authority has required the manufacturer, in accordance with Article 97(1) MDR, to carry out the applicable conformity assessment procedure (may be provided upon request)
  - Expired/*expires after* 20 March 2023:
    - A formal application to the notified body in accordance with Section 4.3, first subparagraph of Annex VII MDR for conformity assessment has been made or will be made/submitted by us to a notified body no later than 26 May 2024 for the device(s) listed in the attached schedule or its substitute and a signed written agreement is/will

---

<sup>1</sup> The first condition is not applicable in case of devices for which the conformity assessment procedure pursuant to MDD did not require the involvement of a notified body, for which the declaration of conformity was drawn up prior to 26 May 2021 and for which the conformity assessment procedure pursuant to this Regulation requires the involvement of a notified body

be in place in accordance with Section 4.3, second subparagraph of Annex VII MDR before 26 September 2024.

➤ **Up-classified devices**

In case of devices for which the conformity assessment procedure pursuant to MDD did not require the involvement of a notified body, for which the declaration of conformity was drawn up prior to 26 May 2021 and for which the conformity assessment procedure pursuant to this Regulation requires the involvement of a notified body:

*Choose one applicable statement:*

- A formal application to the notified body in accordance with Section 4.3, first subparagraph of Annex VII MDR for conformity assessment has been made or will be made/submitted by us to a notified body no later than 26 May 2024 for the device(s) listed in the attached schedule or its substitute and a signed written agreement is/will be in place in accordance with Section 4.3, second subparagraph of Annex VII MDR before 26 September 2024.

➤ **Quality Management System (QMS)**

• *Choose one applicable statement:*

- A QMS in accordance with Article 10(9) MDR will be put in place by no later than 26 May 2024.
- A QMS in accordance with Article 10(9) MDR is in place.
- A notified body has issued the attached certificate for the MDR-compliant QMS.

➤ **Device(s) as listed in the attached schedule**


- The device(s) continue to comply with the AIMDD or MDD.
- The device(s) has/have not been significantly changed in its/their design and intended purpose since 26 May 2021.
- The device(s) do not present an unacceptable risk to health or safety of patients, users, or other persons, or to other aspects of the protection of public health.

**Signed for and on behalf of the manufacturer:**

Full Company Name: Kuraray Noritake Dental Inc.

Location & Date: Tokyo, Japan, 2023-09-05

Signature, Print Name, Title:



Yasujiro Ohara, General Manager of Quality Assurance Department

The above Manufacturer's Declaration is valid for the following devices:

<b>Device identification</b> <i>(e.g., device name, family / group name device model or catalogue number)</i>	<b>Directive Certificate number(s) to which this confirmation is made</b>  <input type="checkbox"/> <i>Not applicable</i>	<b>Original expiry date<sup>2</sup></b>  <input type="checkbox"/> <i>Not applicable</i>	<b>Notified Body name and number</b>  <input type="checkbox"/> <i>Not applicable</i>	<b>Classification and rule under the MDR</b>	<b>End date of extended validity / transition period</b>	<b>Substitute Device</b>  <input checked="" type="checkbox"/> <i>Not applicable</i>
Dental Porcelain	JP19/040488	2023-10-24	SGS Belgium NV  1639	Ila  Rule 8	2028-12-31	N/A
Dental Ceramic Block	JP19/040488	2023-10-24	SGS Belgium NV  1639	Ila  Rule 8	2028-12-31	N/A

---

<sup>2</sup> As indicated on the Directive Certificate prior to the extension of the validity.